


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130303 ✓
Name	JAMES I JENSEN, JR	ID#	3393
Date	11-30-10 ✓		
A	Agency	YAKUTAT DPS	Phone # 784-3206
	Instrument Location	AT POLICE STATION	
B	Alco S/N	77286 ✓	Target Value .082 ✓
			High Pressure 225
	Alco Test Values	.080 ✓	.080 ✓
		1st Alco	2nd Alco
	Signature		
		(OVER)	


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 5th day of JAN, 2011.


 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130303

BMB
12/16/10

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

NOVEMBER 30, 2010

OPERATOR'S NAME:

JENSEN JAMES I

OPERATOR'S NUMBER: 3393

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: YAK1

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: 77286

— BREATH ANALYSIS —

.002 ADJUSTED FOR 29.78 in
ALCO TARGET .001 19:31
BLANK TEST .000 19:32
INTERNAL STANDARD VERIFIED 19:32
ALCO TV 29.78 in .000 19:32
BLANK TEST .000 19:33
SUBJECT SAMPLE .000 19:33
BLANK TEST .000 19:34
ALCO TV 29.78 in .000 19:34
BLANK TEST .000 19:35

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

NOVEMBER 30, 2010

TIME 19:25

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 29.77 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢